

Calibration procedure for the hardware of RNO-G [Time: 10+3]

Thursday, 8 April 2021 10:48 (13 minutes)

Description of the test and calibration procedure for the hardware of the radio neutrino detector RNO-G: the three types of amplifier boards, which will be used in the detector, were tested with a vector network analyzer and in a temperature chamber.

Primary author: CATALDO, Maddalena (DESY Zeuthen)

Presenter: CATALDO, Maddalena (DESY Zeuthen)

Session Classification: Radio